



**HAL**  
open science

## Proficiency of medical students at obtaining pressure measurement readings using Automated ankle and toe measuring devices for diagnosis of lower extremity peripheral artery disease

F Catillon, S Tuffier, A Guilcher, Q Tollenaere, A Métairie, A Miossec, C Mauger, D Laneelle, Guillaume Mahé

### ► To cite this version:

F Catillon, S Tuffier, A Guilcher, Q Tollenaere, A Métairie, et al.. Proficiency of medical students at obtaining pressure measurement readings using Automated ankle and toe measuring devices for diagnosis of lower extremity peripheral artery disease. *Annals of Vascular Surgery*, 2020, 65, pp.183-189. 10.1016/j.avsg.2019.10.092 . hal-02394264

**HAL Id: hal-02394264**

**<https://univ-rennes.hal.science/hal-02394264>**

Submitted on 22 Aug 2022

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

1 **Title:** Proficiency of medical students at obtaining pressure measurement readings using  
2 Automated ankle and toe measuring devices for diagnosis of lower extremity peripheral artery  
3 disease.

4

5 **Short Title:** Medical students and automated pressure devices

6

7 **Authors list :** Catillon F<sup>1</sup>, Tuffier S<sup>2</sup>, Guilcher A<sup>1,3</sup>; Tollenaere Q<sup>1,3</sup>; Métairie A<sup>1,3</sup>, Miossec  
8 A<sup>1,3</sup>; Mauger C<sup>4</sup>, Laneelle D<sup>5</sup>, Mahé G<sup>1,3</sup>.

9

10 **Affiliations :**

11 1 CHU Rennes, Médecine vasculaire France.

12 2 Univ Rennes 1 Public health Department, France

13 3 Univ Rennes 1 ; INSERM CIC 1414, France

14 4 Médecine Vasculaire, CH Saint-Malo, France

15 5 CHU Caen Normandie, Médecine vasculaire, France

16

17

18 **Corresponding author :**

19 Guillaume MAHE.

20 Pôle imagerie médicale et explorations fonctionnelles. Hôpital Pontchaillou, 2 rue Henri Le  
21 Guilloux. Rennes, F-35033. France. Tel: +33 (0)2 9928 4321. Fax: +33 (0)2 9928 4364

22 E-mail: maheguillaume@yahoo.fr

23

24 Twitter : @GMahe\_

25

26

27

28 **ABSTRACT:**

29

30 **OBJECTIVE:** Pressure measurement is a key component in the diagnosis of lower extremity  
31 peripheral artery disease (PAD) but is technically challenging and time-consuming for non-  
32 vascular specialists, thus hindering its wider implementation. The aim of this study was to  
33 assess the proficiency of students at obtaining satisfactory ankle or toe pressure readings for  
34 PAD diagnosis using two automated devices.

35 **METHODS:** Medical students followed a training session after which they performed ankle  
36 and toe pressure measurements to calculate the ankle-brachial index (ABI) using the MESI  
37 ABP MD® device, and the toe-brachial index (TBI) using the SYSTOE® device. Blinded  
38 vascular specialists took the same measurements. Use of the automated devices was  
39 considered satisfactory when a valid reading was measured in as few attempts as possible. A  
40 comparison was made of each student's proficiency at performing valid ankle and toe  
41 pressure measurements. The secondary objective was to compare the readings taken by the  
42 vascular specialists with those of the students.

43 **RESULTS:** Forty-three medical students were included. Mean number of attempts was 1.23  
44 +/- 0.48 with the MESI ABP MD® device and 1.44 +/- 0.55 with the SYSTOE® device (p =  
45 0.04). There was no statistically significant difference between ABI readings taken by the  
46 students and those taken by the vascular specialists, 1.17 [0.90; 1.39 ] versus 1.18 [0.86; 1.39  
47 ] (p = 0.33), contrary to TBI readings 0.70 [0.22; 1.74] versus 0.72[0.23; 1.16] (p = 0.03).  
48 Measurement duration for the students and vascular specialists was 3.75 min±1.12 and 2.26  
49 min ±0.82 (p < 0.01) with the MESI ABP MD® device and 4.30 min ±1.23 and 3.33 min  
50 ±1.49 (p = 0.03) with the SYSTOE® device. Correlation coefficients between the students  
51 and the vascular specialists were 0.56 and 0.34 with the MESI ABP MD® and SYSTOE®  
52 devices (p < 0.05).

53 **CONCLUSION:** After a brief theoretical training session, the medical students were better at  
54 taking ankle pressure measurements than toe pressure measurements with an automated  
55 device for the purposes of PAD diagnosis. It would be of value to assess the advantages of  
56 these automated devices in primary care practice in future research.

57

58

## 59 INTRODUCTION

60 Lower extremity peripheral artery disease (PAD) is a common medical condition affecting  
61 almost 202 million people worldwide(1). The European Society of Cardiology (ESC) and the  
62 American Heart Association (AHA) recommend ankle-brachial index (ABI) pressure  
63 measurement as a first-line non-invasive examination technique to investigate and establish  
64 diagnosis of PAD(2,3).

65 Although pressure measurement plays a major role in establishing diagnosis of PAD, little  
66 research has focused on how this measurement technique is taught(4) and the knowledge of  
67 French medical students on this subject leaves much to be desired(5,6) .

68 Wider implementation of this technique however is hampered, especially among non-vascular  
69 physicians, by its high requirements in terms of technical and learning skills. A previous study  
70 revealed that even after following a practical training course, only 11 of 20 students  
71 succeeded in measuring systolic pressure and even then had lost their proficiency six months  
72 later through lack of practice in the intervening period(7,8).

73 In terms of screening, follow-up and overall coordination primary care physicians are key  
74 players in the management of PAD, which represents a serious public health concern with  
75 regard to morbidity and mortality rates and financial impact(9,10). And yet wider  
76 implementation of ABI measurement in primary care settings is severely hindered by time-  
77 restricted primary care consultations, insufficient clinical criteria for the diagnosis of  
78 PAD(11), and the technical and practical skills required to ensure reliable continuous-wave  
79 Doppler readings.

80 Several automated methods for the detection of PAD have been discussed(12), with reference  
81 to certain factors such as diagnostic thresholds or particular populations (for example patients  
82 with diabetes) that are liable to affect their validity(13,14). These methods would, however,  
83 make it much quicker and easier for non-specialists to measure systolic pressure and would  
84 open up new prospects for wider implementation of screening for PAD. This is especially true  
85 given that the training methods for these techniques are potentially far simpler than those of  
86 previously validated systolic pressure techniques.

87 Moreover, medical students as budding physicians are an ideal population to target for greater  
88 awareness of ABI measurement and its training programs(15). On another point, training  
89 resources for these automated methods have yet to be investigated.

90 The main aim of the present study was to assess the proficiency of medical students at  
91 obtaining satisfactory pressure readings (i.e. valid reading obtained in as few attempts as  
92 possible) for diagnosis of PAD using two automated devices: the MESI ABP MD® and the  
93 SYSTOE®. Secondary aims were to assess i) measurement duration, and ii) concordance of  
94 readings taken by the vascular specialists with those taken by the students with the various  
95 devices  
96

## 97 MATERIALS AND METHODS

### 98 Study Design

99 This was a single center controlled prospective study conducted at the Vascular Medicine  
100 Unit of the university hospital of Rennes in France from December 2018 to February 2019. It  
101 was approved by the ethics committee of Rennes Hospital (no. 16.150). Second to 6th year  
102 medical students with no prior training or knowledge of the devices involved were invited by  
103 electronic mail to participate in the trial having first provided written consent. Patients with  
104 scheduled Doppler ultrasound appointments were included and oral consent was provided.  
105 Patients with wounds were excluded.

### 106 Initial training session

107 The initial hour-long training session, dispensed by a vascular specialist, was composed of an  
108 introduction to the pathophysiology of PAD, an update on PAD guidelines and introduction to  
109 standard Doppler ABI measurement. The students then viewed a video presentation of how to  
110 use each device. The MESI ABP MD® simultaneously records systolic ankle and brachial  
111 pressures using three color-coded cuffs (one on the arm and the other two on the ankles with  
112 the tubes pointing upwards)(16). Systolic pressure is calculated within minutes and does not  
113 require pre-measurement resting. The SYSTOE® employs photoplethysmography to record  
114 systolic pressure in the hallux (TBI) after first taking the temperature of the toe with an infra-  
115 red thermometer(16,17). Indeed, when toe temperature is low, the measured pressure may be  
116 falsely low(18). A cuff is placed at the base of the toe with the tube pointing downwards. The  
117 sensor is positioned at the top of the toe by means of an adhesive pad and pressure is applied  
118 to the sensor by the operator during deflation then released. The diagnostic threshold adopted  
119 for PAD was 0.70(19) regarding TBI and 0.90 regarding ABI(20). In the absence of any  
120 research validating the SYSTOE® threshold, a value of 0.70 was adopted.

### 121 Real-life patient assessment:

122 Each student was called individually to measure patient's pressure with both devices within  
123 one week to one month of the initial training session.

124 The procedure consisted of obtaining one pressure reading with the MESI ABP MD® and  
125 another with the SYSTOE® from each foot. Brief written instructions were provided to  
126 facilitate the smooth running of the procedure. The stopwatch used to time measurement  
127 acquisition was started once the first cuff had been put in place. The SYSTOE® was used in

128 semi-automatic mode (in view of the fact that automatic mode requires applying two cuffs  
129 and this can be problematic depending on the specific features of certain toes) and the student  
130 was responsible for determining the inflection point of the curve. Arterial pressure was  
131 measured by an automatic device(21).

132 A blinded vascular specialist present in the unit then took measurements following the same  
133 procedure and concluded by taking Doppler pressure measurements in line with standard  
134 procedure (i.e. right brachial artery, right posterior tibial artery, right anterior tibial artery, left  
135 posterior tibial artery, left anterior tibial artery, left brachial artery and right brachial  
136 artery)(20,22). The highest brachial and ankle pressure values were used to calculate the ABI.

137

138 Patient characteristics:

139 During the 10 minute pre-measurement rest period, the following information was gathered:  
140 cardiovascular risk factors, previous history of kidney failure or cardiovascular disease,  
141 anticoagulant treatment (vitamin K antagonists or direct oral anticoagulants), weight and  
142 height(23).

143

144 Statistical Analysis:

145 Descriptive analysis was used to investigate each variable collected. Quantitative variables  
146 were expressed as mean, standard deviation, median, and maximal and minimal values.  
147 Qualitative variables were expressed as counts and percentages. Systolic pressure readings  
148 from the various devices and time taken by each operator were compared using the Student's  
149 t-test for paired data that followed a Gaussian distribution and the Wilcoxon signed-rank test  
150 for paired data where this was the opposite. A linear mixed effects model was used to assess  
151 time spent on each device and operator status. Measurement concordance was summarized  
152 using the Pearson correlation coefficient, the Bland-Altman plot(24) and Lin's concordance  
153 correlation coefficient(25). The latter simultaneously calculates correlation with and deviation  
154 from perfect agreement between the two measuring devices. Statistical analysis was  
155 conducted using R software (version 3.5.3) and the Tidyverse packages (R core Team 2019,  
156 Hadley Wickham 2017). A p-value of  $p < 0.05$  was regarded as statistically significant.

157

158 RESULTS

159 Forty-three medical students participated in measuring systolic pressure (37.2% 2nd year  
160 students, 20.9% 3rd year students, 16.3% 4th year students, 7.0% 5th year students and 18.6%  
161 6th year students).

162 Forty-three patients (67% male) were included. Their characteristics are shown in Table 1.

163 Prevalence of lower extremity PAD in the 43 subjects as measured by the MESI ABP MD®,  
164 the SYSTOE® and Doppler ultrasound was 4.6 %, 55.0 % et 11.0% respectively (according  
165 to vascular specialist data).

166

167 Proficiency of students at obtaining satisfactory ABI and TBI readings:

168 Mean number of attempts was  $1.23 \pm 0.48$  regarding the MESI ABP MD® and  $1.44 \pm 0.55$  ( $p$   
169  $= 0.04$ ) regarding the SYSTOE® (Figure 1). Forty students (93%) used the MESI ABP MD®  
170 correctly and 31 (72%) used the SYSTOE® correctly ( $p = 0.28$ ). Errors in relation to the  
171 MESI ABP MD® occurred due to incorrect placement (such as cuff tubes pointing in the  
172 wrong direction, incorrect cuff placement or incorrect cuff size). As for the SYSTOE®  
173 device, problems essentially arose from the positioning of the sensor and cuffs applied back to  
174 front.

175 Ninety-three percent of the students took toe temperature before using the SYSTOE®. The  
176 MESI ABP MD® screen displayed a total of 8 readings indicating PAD and 2 messages  
177 indicating inflation error (error E2).

178

179

180 Comparison of measurement duration on each device:

181 The different duration of the measurements are presented in figure 2. As determined by a  
182 multiple regression model, MESI ABP MD® measurement took 3.7 times less time than  
183 Doppler ultrasound ( $-3.713 \pm 0.170$ ) and SYSTOE® measurement took 2.9 times less time ( $-$   
184  $2.913 \pm 0.171$ ), demonstrating statistical significance. The vascular specialists on the whole  
185 took 1.2 times less time than the students irrespective of the device used ( $-1.238 \pm 0.130$ ).

186

187 Comparison of systolic pressure readings



188 No statistically significant difference was found between ABI readings obtained by the  
189 students and those obtained by the vascular specialists with the MESI ABP MD®: 1.17 [0.90;  
190 1.39] versus 1.18 [0.86; 1.39] ( $p = 0.33$ ), unlike TBI readings with the SYSTOE®: 0.70 [0.22;  
191 1.74] versus 0.72[0.23; 1.16] ( $p = 0.03$ ). There were 20 missing ABI values in student data  
192 (23.3%) versus 22 (25.6%) in specialist data and 14 missing TBI values (16.3%) versus 15  
193 (17.4%).

194 The correlation coefficient ( $r$ ) between the readings obtained by the students and those  
195 obtained by the vascular specialists was 0.56 using the MESI ABP MD® and 0.34 using the  
196 SYSTOE®  $p < 0.05$ . The Lin concordance correlation coefficient was 0.55 [0.36; 0.70] and  
197 0.34 [0.12; 0.53] respectively. Correlation between systolic pressure readings obtained by the  
198 specialists and students per device is shown in Figure 3. With respect to the vascular  
199 specialists, the correlation coefficient between the MESI ABP MD® and Doppler ultrasound  
200 was 0.2 ( $p = 0.11$ ) and the Lin concordance correlation coefficient was 0.18 [-0.04; 0.39].

201

202

## 203 DISCUSSION

204 This is the first research project to focus on the efficacy of training in ABI and TBI  
205 measurement using automated measuring devices and its subsequent implementation by  
206 medical students in the diagnosis of PAD.

207 This study highlights the fact that the learning benefits of a single training session are not the  
208 same from one automated measuring method to the next. For example, the MESI ABP MD®  
209 system entailed less attempts at obtaining a satisfactory reading and was found to be more  
210 user-friendly than the SYSTOE®. Thus, a single hour-long session combined with a video  
211 presentation is not enough to acquire the skills required for toe measurement, unlike the MESI  
212 ABP MD®. Donnou et al. illustrated that students failed to obtain systolic pressure readings  
213 using continuous-wave Doppler ultrasound after a single theoretical session combined with a  
214 video presentation(8). Mahe demonstrated that the teaching of systolic pressure measurement  
215 in medical faculties was predominantly based on theoretical training(26). Lecturers in  
216 medicine keen for students to learn how to obtain reliable pressure readings should introduce  
217 the measuring devices into their theoretical training sessions. Automated measuring devices  
218 have greater benefits than standard Doppler techniques, whose measurement duration

219 produced statistically lower values irrespective of the operator. The vascular specialists,  
220 however, outperformed the students in terms of measurement duration, emphasizing that  
221 practice is required. This is an important point in the light of increasing patient numbers and  
222 time-restricted consultations in general practice (mean time of 16.4 minutes)(27,28). And yet  
223 it is precisely during these consultations that patients should have access to this type of  
224 screening. Moreover, correlation between the students and the specialists was only 0.56 with  
225 the MESI ABP MD® and 0.34 with the SYSTOE®. This may result from lack of experience,  
226 especially when using the SYSTOE® which requires a certain level of technical skill to  
227 obtain a reading and identify the point of inflection that determines the pressure value. This is  
228 moreover corroborated by Bland-Altman results that were indicative of lower measurement  
229 variability between the students and the vascular specialists with the MESI ABP MD®.

230 A further point of importance aside from learning is the reliability of the readings obtained for  
231 the diagnosis of PAD. The MESI ABP MD® and the SYSTOE® have already been endorsed  
232 in clinical trials over corresponding gold standard measurement methods, namely continuous-  
233 wave Doppler ABI measurement and laser Doppler velocimetry respectively(15,16). In the  
234 present study, only the performance of the MESI ABP MD® was compared with continuous-  
235 wave Doppler ultrasound, yielding an extremely low Lin concordance coefficient of 0.18.  
236 One possible explanation is that MESI ABP MD® measurement was done first and Doppler  
237 measurement last. There was therefore a time lapse of at least 15 minutes between the two  
238 readings. Furthermore, it is important to note that our population had no links with PAD  
239 screening program but was a pre-selected, hospital population. In fact the present study  
240 accentuates the importance of selecting a method of diagnosis adapted to the population in  
241 question. The discrepancy found between PAD prevalence determined by the SYSTOE® and  
242 that determined by Doppler ultrasound may be attributed to the participation of 11 diabetic  
243 patients and 16 renal insufficiency patients (of whom 2 were patients with both diabetes and  
244 renal failure), meaning 25 patients (58%) from our population were potentially at risk of  
245 increased arterial stiffness. Hence the possibility of falsely normal systolic pressure readings  
246 in these patients.

247

248 The present study also indicated a slight upward bias in the systolic pressure readings  
249 obtained from the MESI ABP MD® versus Doppler ultrasound (mean ABI 1.18+/-0.12  
250 versus 1.15+/-0.20). These results are consistent with those found in the literature and may

251 stem from observer bias where Doppler ultrasound is concerned, such as the time required to  
252 perceive the signal and stop deflating the cuff(18,28). This raises the issue of whether lower  
253 extremity PAD detection thresholds using these automated tools should be redefined, as  
254 previously suggested by other authors(16). A meta-analysis by Verberk *et al.*(29) expresses a  
255 preference for a cut-off in the region of 1 for the purposes of improving the PAD detection  
256 sensitivity in these devices.

257 Unlike Doppler ultrasound, the MESI ABP MD® provides no precise details on the artery  
258 used for recording, and placing the cuff on the right arm may be questionable in the event of  
259 subclavian artery stenosis, whose prevalence is estimated to be between 2% and 5% in the  
260 general population(30). The scope of both the MESI ABP MD® and continuous-wave  
261 Doppler ultrasound is also limited in certain patient groups (such as diabetic, elderly or renal  
262 failure patients) where readings may be falsely reassuring, contrary to the SYSTOE® whose  
263 measuring technique is unaffected by increased arterial stiffness. This is exemplified by  
264 disparity in prevalence from one measuring tool to another.

#### 265 Limitations

266 This study has several limitations. Firstly, each system involves different pressure  
267 measurement sites. Toe-brachial pressure is undoubtedly more difficult to measure for the  
268 inexperienced, even with an automated device. Our aim was to assess the training methods for  
269 these automated tools. As such, it was found that following an hour-long training session, the  
270 students were more proficient at measuring ABI. A further limitation is that the participating  
271 students all came from the same faculty of medicine, although we made very sure that they  
272 were novices in measurement to avoid any previous training interfering with the results. The  
273 level of study of the student cohort taking the measurements was diverse and it was  
274 impossible to assess any influence of this diversity on proficiency at taking measurements. On  
275 another note, it was unfeasible to randomize the order of assessment of the measuring tools. It  
276 is impossible to rule out any potential impact of this factor on the results. However, the  
277 students were tested first on the easier MESI ABP MD®, conceivably making them feel more  
278 confident about measuring with the SYSTOE® second. A final limitation accounting for  
279 variation in the readings was inter-operator variability regarding the vascular specialists when  
280 measuring ABI and TBI.

281

282 CONCLUSION

283 Following an hour-long theoretical training session, the medical students were more proficient  
284 at obtaining ABI readings using an automated device than at obtaining TBI readings and were  
285 able to execute both far more quickly than with the gold standard Doppler ultrasound method  
286 used by the vascular specialists. This research project has demonstrated that even where  
287 automated devices are concerned it is of fundamental necessity to adopt specific training  
288 methods for pressure measurement, without disregarding the importance of practical  
289 experience. The benefits of implementing routine use of these tools in primary care practice  
290 have yet to be assessed.

291

292 Acknowledgement: The authors thank the MESI company which loan the MESI Device. The  
293 MESI company had no role in the design and interpretation of the results of the study.

294

295 Funding: This research did not receive any specific grant from funding agencies in the public,  
296 commercial, or not-for-profit sectors.

297

298 Conflict of interest: None

299

300

301

302

303

304

305

306

307

308

309

310

311

312

313

314

315 REFERENCES:

- 316 1. Fowkes FGR, Rudan D, Rudan I, Aboyans V, Denenberg JO, McDermott MM, et al.  
317 Comparison of global estimates of prevalence and risk factors for peripheral artery  
318 disease in 2000 and 2010: a systematic review and analysis. *Lancet*. 2013 Oct  
319 19;382(9901):1329–40.
- 320 2. Gerhard-Herman MD, Gornik HL, Barrett C, Barshes NR, Corriere MA, Drachman DE,  
321 et al. 2016 AHA/ACC Guideline on the Management of Patients With Lower Extremity  
322 Peripheral Artery Disease: A Report of the American College of Cardiology/American  
323 Heart Association Task Force on Clinical Practice Guidelines. *Circulation*. 2017 Mar  
324 21;135(12):e726–79.
- 325 3. Aboyans V, Ricco J-B, Bartelink M-LEL, Björck M, Brodmann M, Cohnert T, et al.  
326 2017 ESC Guidelines on the Diagnosis and Treatment of Peripheral Arterial Diseases, in  
327 collaboration with the European Society for Vascular Surgery (ESVS): Document  
328 covering atherosclerotic disease of extracranial carotid and vertebral, mesenteric, renal,  
329 upper and lower extremity arteries Endorsed by: the European Stroke Organization  
330 (ESO) The Task Force for the Diagnosis and Treatment of Peripheral Arterial Diseases of  
331 the European Society of Cardiology (ESC) and of the European Society for Vascular  
332 Surgery (ESVS). *Eur Heart J*. 2018 Mar 1;39(9):763–816.
- 333 4. Omarjee L, Donnou C, Chaudru S, Locher C, Paul E, Charasson M, et al. Impact of an  
334 Educational Intervention on Ankle-Brachial Index Performance Among Medical  
335 Students and Fidelity Assessment at Six Months. *Ann Vasc Surg*. 2019 Apr;56:246–53.
- 336 5. Stivalet O, Omarjee L, Chaudru S, Hoffmann C, Bressollette L, Cohoon KP, et al.  
337 Noninvasive Peripheral Artery Disease Screening Tools: A Deficient Knowledge among  
338 French Vascular Residents from 4 Medical Schools. *Ann Vasc Surg*. 2018 Feb;47:134–  
339 42.
- 340 6. Lanéelle D, Hoffmann C, Stivalet O, Omarjee L, Mahé G. Vascular medicine residents  
341 lack adequate training for limb pressure measurement: A nationwide survey in France.  
342 *Vasc Med*. 2019 Oct;24(5):452–4.
- 343 7. Donnou C, Chaudru S, Stivalet O, Paul E, Charasson M, Selli J-M, et al. Medical  
344 students' proficiency in performance of the resting ankle-brachial index is not sustained  
345 at 6 months. *J Clin Hypertens (Greenwich)*. 2018 Mar;20(3):609–10.
- 346 8. Donnou C, Chaudru S, Stivalet O, Paul E, Charasson M, Selli J-M, et al. How to become  
347 proficient in performance of the resting ankle-brachial index: Results of the first  
348 randomized controlled trial. *Vasc Med*. 2017 Nov 1;1358863X17740993.

- 349 9. Diehm C, Allenberg JR, Pittrow D, Mahn M, Tepohl G, Haberl RL, et al. Mortality and  
350 vascular morbidity in older adults with asymptomatic versus symptomatic peripheral  
351 artery disease. *Circulation*. 2009 Nov 24;120(21):2053–61.
- 352 10. Hirsch AT, Hartman L, Town RJ, Virnig BA. National health care costs of peripheral  
353 arterial disease in the Medicare population. *Vasc Med*. 2008 Aug;13(3):209–15.
- 354 11. Khan NA, Rahim SA, Anand SS, Simel DL, Panju A. Does the clinical examination  
355 predict lower extremity peripheral arterial disease? *JAMA*. 2006 Feb 1;295(5):536–46.
- 356 12. Wohlfahrt P, Ingrischová M, Krajcoviechová A, Palous D, Dolejšová M, Seidlerová J, et  
357 al. A novel oscillometric device for peripheral arterial disease screening in everyday  
358 practice. The Czech-post MONICA study. *Int Angiol*. 2011 Jun;30(3):256–61.
- 359 13. Herráiz-Adillo Á, Cavero-Redondo I, Álvarez-Bueno C, Martínez-Vizcaíno V, Pozuelo-  
360 Carrascosa DP, Notario-Pacheco B. Factors affecting the validity of the oscillometric  
361 Ankle Brachial Index to detect peripheral arterial disease. *Int Angiol*. 2017  
362 Dec;36(6):536–44.
- 363 14. Mayr V, Hirschl M, Klein-Weigel P, Girardi L, Kundi M. A randomized cross-over trial  
364 in patients suspected of PAD on diagnostic accuracy of ankle-brachial index by Doppler-  
365 based versus four-point oscillometry based measurements. *VASA*. 2019 Jul 5;1–7.
- 366 15. Chaudru S, de Müllenheim P-Y, Le Faucheur A, Jaquinandi V, Mahé G. Ankle brachial  
367 index teaching: A call for an international action. *Int J Cardiol*. 2015 Apr 1;184:489–91.
- 368 16. Špan M, Geršak G, Millasseau SC, et al. Detection of peripheral arterial disease with an  
369 improved automated device: comparison of a new oscillometric device and the standard  
370 Doppler method. *Vasc Health Risk Manag*. 12:305–11.
- 371 17. Pérez-Martin A, Meyer G, Demattei C, Böge G, Laroche J-P, Quéré I, et al. Validation  
372 of a fully automatic photoplethysmographic device for toe blood pressure measurement.  
373 *Eur J Vasc Endovasc Surg*. 2010 Oct;40(4):515–20.
- 374 18. Sawka AM, Carter SA. Effect of temperature on digital systolic pressures in lower limb  
375 in arterial disease. *Circulation*. 1992 Mar;85(3):1097–101.
- 376 19. Høyer C, Sandermann J, Petersen LJ. The toe-brachial index in the diagnosis of  
377 peripheral arterial disease. *J Vasc Surg*. 2013 Jul;58(1):231–8.
- 378 20. Aboyans V, Criqui MH, Abraham P, Allison MA, Creager MA, Diehm C, et al.  
379 Measurement and interpretation of the ankle-brachial index: a scientific statement from  
380 the American Heart Association. *Circulation*. 2012 Dec 11;126(24):2890–909.
- 381 21. Stivalet O, Paisant A, Belabbas D, Omarjee L, Le Faucheur A, Landreau P, et al.  
382 Exercise testing criteria to diagnose lower extremity peripheral artery disease assessed  
383 by computed-tomography angiography. *PLoS ONE*. 2019;14(6):e0219082.
- 384 22. Chaudru S, de Müllenheim P-Y, Le Faucheur A, Kaladji A, Jaquinandi V, Mahé G.  
385 Training to Perform Ankle-Brachial Index: Systematic Review and Perspectives to  
386 Improve Teaching and Learning. *Eur J Vasc Endovasc Surg*. 2016 Feb;51(2):240–7.

- 387 23. Sadler SG, Hawke FE, Chuter VH. The effect of pretest rest time on automated measures  
388 of toe systolic blood pressure and the toe brachial index. *Blood Press Monit.* 2015  
389 Oct;20(5):245–8.
- 390 24. Martin Bland J, Altman DouglasG. Statistical methods for assessing agreement between  
391 two methods of clinical measurement. *The Lancet.* 1986 Feb 8;327(8476):307–10.
- 392 25. Lin LI-K. A Concordance Correlation Coefficient to Evaluate Reproducibility.  
393 *Biometrics.* 1989;45(1):255–68.
- 394 26. Mahé G. [Ankle-brachial index measurement: Methods of teaching in French medical  
395 schools and review of literature]. *J Mal Vasc.* 2015 May;40(3):165–72.
- 396 27. Deveugele M, Derese A, van den Brink-Muinen A, Bensing J, De Maeseneer J.  
397 Consultation length in general practice: cross sectional study in six European countries.  
398 *BMJ.* 2002 Aug 31;325(7362):472.
- 399 28. Iacobucci G. GP appointments last less than five minutes for half the world’s population.  
400 *BMJ.* 2017 Nov 8;359:j5172.
- 401 29. Verberk WJ, Kollias A, Stergiou GS. Automated oscillometric determination of the  
402 ankle-brachial index: a systematic review and meta-analysis. *Hypertens Res.* 2012  
403 Sep;35(9):883–91.
- 404 30. Shadman R, Criqui MH, Bundens WP, Fronck A, Denenberg JO, Gamst AC, et al.  
405 Subclavian artery stenosis: prevalence, risk factors, and association with cardiovascular  
406 diseases. *J Am Coll Cardiol.* 2004 Aug 4;44(3):618–23.

407

408

409

410

411


412 Figure 1 Title: Proficiency of students at obtaining satisfactory Ankle brachial index (MESI  
413 device) and Toe Brachial index (SYSTOE) readings

414


415 Figure 2 Title: Comparison of measurement durations on each device.

416

417 Figure 3 Title: Bland and Altman representation between systolic pressure readings obtained  
418 by the specialists and students per device.











Table 1: Patients' characteristics

Clinical characteristics	n=43
Men, n (%)	29 (67)
Age (Years), mean ± standard deviation	66 ±14,4
Body mass index (Kg/m <sup>2</sup> ) mean ± standard deviation	26,9 ±4,9
Hypertension, n (%)	30 (70)
Tobacco, n (%)	4 (9)
Diabetes, n (%)	11 (25)
Dyslipidemia, n (%)	18 (42)
Renal insufficiency, n (%)	16 (37)
Cardiovascular diseases, n (%)	20 (46)
Anticoagulants, n (%)	8 (19)

Table 1 legend: Hypertension was defined as patients receiving antihypertensive drugs or an arterial pressure >140/90 mmHg. Dyslipidemia was defined as patients receiving statins. Diabetes was defined as patients receiving antidiabetic drugs. Renal insufficiency was defined as creatinine clearance <60ml/min. Cardiovascular diseases was defined as a history of coronary artery disease or peripheral artery disease.